

**Tape Cartridge part# C190T-U**

[ Please scan both barcodes for your selection]



Measurement interval 1/10 inch



Measurement interval 1/4 inch



Measurement interval 1/2 inch




Measurement interval inch nearest



Measurement interval inch round up

**Calibration**

To calibrate the device after replacing the tape cartridge, scan both barcodes, extract tape to 30cm and press measure. This reconfirms the zero.



**Tape Cartridge part# C190T-RP C190T-S**

[ Please scan both barcodes for your selection]



C190T-RP Nearest inch



C190T-S Inch round up

**Calibration**

To calibrate the device after replacing the tape cartridge, scan both barcodes, extract tape to 30cm and press measure. This reconfirms the zero.

